


# Semimaschera facciale FFP2 NR

Standard: EN149: 2001+A1: 2009



- Bassa resistenza respiratoria
- 4 Strati
- FFP2 con livello di protezione ad alta efficienza



TNT in PP

Filtro soffiato a fusione in PP

Filtro di cotone ad aria calda

TNT in PP



# Informazioni per gli ordini

Nome del prodotto
SEMIMASCHERA FACCIALE FILTRANTE
FFP2 NR III SENZA VALVOLA MONOUSO
NON STERILE

## Codice prodotto KF-2


Confezione di vendita:


1 Scatola da 20 pezzi in busta singola





# KF-2 Montaggio Semimaschera FFP2 NR







## CE CERTIFICATE



## Module B EU Type-Examination Certificate

For the requirements of PPE Regulation 2016/425

Certificate No.: CE-PC-200526-433-01-9B

Certificate B.J.ZH.F.PANTHER MEDICAL EQUIPMENT CO., LTD.

holder: Floor 3, Building 1, 28 Huoju Street, Changping Science and

Technology Park, Changping District, 102200 Beijing, China

Product: Particle Filtering Half Mask

Detailed product description listed in the Annex

Model(s): KF-2

Standard(s): EN 149:2001+A1:2009

Respiratory protective devices - Filtering half masks to protect against

particles - Requirements, testing, marking

Issue date: 2020-06-15

Revision date: 2020-08-10

Expiry date: 2021-06-14

The product(s) on this certificate and the Technical File have been assessed and found to be in conformance with the applicable Essential Health and Safety Requirements in Annex II of the PPE regulation 2016/425.

Any changes to the design, manufacturing location or manufacture of the PPE product certified here must be advised to CCQS Certification Services Limited for review.

CE marking shall not be applied until the requirements of all the PPE Regulation 2016/425 and relevant EN Harmonised standards and/or Technical specifications have been met.

If the certified product is Category III then this certificate is only valid if used in conjunction with Conformity Assessment against Module C2 or Module D.

This certificate remains the property of CCQS and maybe withdrawn at any time if it is considered that the equipment is no longer in conformity with the requirements of the PPE Regulation 2016/425.



Approved by Ireland Government as a Notified Body for CE Marking No.2834





#### **CCQS Certification Services Limited**

Block 1 Blanchardstown Corporate Park, Ballycoolin Road, Blanchardstown, Dublin15, D15 AKK1, Ireland



## CE CERTIFICATE



# Module B EU Type-Examination Certificate Annex

For the requirements of PPE Regulation 2016/425

Certificate No.: CE-PC-200526-433-01-9B

#### Applicable standards and specification:

EN 149:2001+A1:2009 Respiratory protective devices - Filtering half masks to protect against particles - Requirements, testing, marking

Model reference	Product description
KF-2	Folding filtering half mask fitted with ear loops with head harness clip, no valves, internal metal nose clip
	Classification: FFP2 NR
	Test report No.: 2020(D) - 0732

Certificate Revision	Revision date	Revision details
A	2020-06-15	S Irox Initial issue
В	2020-08-10	Certificate validity extended to one year





## CE CERTIFICATE



## Certificate of Module C2 production monitoring for equipment within the scope of Personal Protective Equipment Regulation (EU) 2016/425 Category III

FPC Certificate No.: CE-PC-200717-556-FPC-B

Certificate B.J.ZH.F.PANTHER MEDICAL EQUIPMENT CO., LTD.

holder: Floor 3, Building 1, 28 Huoju Street, Changping Science and Technology Park, Changping District, 102200 Beijing, China

Manufacturing Floor 3, Building 1, 28 Huoju Street, Changping Science and Technology Park, Changping District, 102200 Beijing, China

Technology Park, Changping District, 102200 Beijing, China Westside No.2, White Bridge Industrial Area, Jinzhan Street,

Chaoyang District ,100018 Beijing, China

1-A, 1-06, 1-07, 1-08, 1-11, 1-12, 1-13, 2-A, 2-12, 3-A, No. 8, Shangcheng Road, Beiqijia Town, Changping District, 102209

Beijing, China

The scope of the The manufacture of respiratory protective device

certification for: See annex for articles covered by this certificate

Validity from: 2020-06-15

Revision date: 2020-08-10

To: 2021-06-14

CCQS Certification Services Limited in its role as a Notified Body for PPE Regulation, is monitoring that the manufacturer is producing PPE in conformity with the type described in the EU type-examination certificate and associated technical file and which satisfies the Essential Health and Safety Requirements of the Regulation. The equipment covered by this certificate is listed in the accompanying schedule. This certificate is not complete and has no validity without the accompanying schedule and revision index.

The manufacturer is hereby authorized to affix our Notified Body number, 2834, to each item of PPE mentioned in the schedule which accompanies this certificate whilst this certificate remains valid.

This certificate and the accompanying schedule remain the property of CCQS and maybe withdrawn or revised at any time if CCQS considers that the equipment is no longer in conformity with the requirements of the Regulation.



Approved by Ireland Government as a Notified Body for CE Marking No.2834





#### **CCQS Certification Services Limited**

Block 1 Blanchardstown Corporate Park, Ballycoolin Road, Blanchardstown, Dublin15, D15 AKK1, Ireland



## **EU Declaration of Conformity**

### **EU Declaration of Conformity**

Annex IX PPE Regulation (EU) 2016/425

This EU Declaration of conformity refers to the following products

(1)	Product Name	Model	Classification/Type	Batch No./Serial No./Identifier
	Particle Filtering half mask	KF-2	FFP2 NR	жжжж

(2) The Manufacturer's name and address is as follows:

Name:	B.J.ZH.F.PANTHER MEDICAL EQUIPMENT CO., LTD				
Address:	Floor 3, Building 1, 28 Huoju Street, Changping Science and Technology Park, Changping District, 102200 Beijing, China.				
	Westside No.2, White Bridge Industrial Area, Jinzhan Street, Chaoyang District, 100018 Beijing, China.				

- (3) This Declaration of Conformity is issued under the sole responsibility of the Manufacturer.
- (4) Detailed description of the PPE to allow traceability/identification of the PPE.

Model: KF-2, Classification: FFP2 NR, White folder half mask without valve









- (5) The article identified in (4) above is in conformance with the relevant Union Harmonization Legislation Regulation (EU) 2016/425.
- (6) References to the relevant harmonized standards used, including the date of the standard, or references to the other technical specifications, including the date of the specification, in relation to which conformity is declared:

No.		Harmonized standard name	
1	EN 149:2001+A1:2009		

(7) CCQS Certification Services Limited. (NB 2834) performed the EU Type Examination (Module B) and issued the Type Examination Certificate Number:

52 54	No.	EU Type Examination (Module B) Certificate Number
	1	CE-PC-200526-433-01-9A

(8) Product Category:

ш	This	prod	luct	IS	Category	11.	-
---	------	------	------	----	----------	-----	---

This product is Category III and is subject to Module C2 internal production control plus supervised product checks at random intervals and is under the surveillance of CCQS Certification Services Limited. (NB 2834)

☐ This product is Category III and is subject to Module D Conformity to type based on quality assurance of the production process and is under the surveillance of CCQS Certification Services Limited. (NB 2834)

Signature: VICky Lee Date: 2020.06.15 Company stamp and/or legal signature:

主直通尔特诺的科性股份有限公司



## Test report: EN149:2001 +AI:2009





中国认可 国际互认 检测 TESTING CNAS L1499 National Quality Supervision and Testing Center for Personal Protective Equipment (Beijing) (Testing Laboratory for Labour Protection Products of Beijing Municipal Institute for Labour Protection)

No.55 Taoranting Street, Xicheng District, Beijing, China. Phone: +86 10 63519250 +86 10 63520770 +86 10 83530311 Fax: +86 10 63519250 +86 10 63520770

The Testing Center is accredited for compliance with ISO/IEC 17025.

The results of tests, calibrations and/or measurements included in this document are traceable to Chinese/national standards.

CNAS is a signatory to the ILAC mutual recognition arrangement for the mutual recognition of the equivalence of testing, calibration and inspection reports.

#### TEST REPORT

Particulate respirator-half facepiece

EN 149: 2001 +A1: 2009 Respiratory protective devices — Filtering half masks to protect against particles — Requirements, testing, marking

Product: Particle filtering half mask

Report No: 2020 (D) - 0732

Client: B. J. ZH. F. PANTHER MEDICAL EQUIPMENT CO., LTD

Model (s): KF-2

Date(s) of tests: 2020.05.15-2020.05.29

DESCRIPTION OF SAMPLES

FFP2 NR

General Information

Classification FFP2 NR. Main Components White folding mask

Manufacturer B. J. ZH. F. PANTHER MEDICAL EQUIPMENT CO., LTD

Manufacturer Address Floor 3, Building 1, 28 Huoju Street, Changping Science and Technology Park, Changping

District, 102200 Beijing, China.

Signed:

Issued: 2020.5.29

陈倬为 Chen Zhuowei Authorized Signatory, Lab Director

Page 1 of 10

This report may not be published except in full unless permission for the publication on approved extract has been obtained in writing.

8 6 8 6 C P A A E E E E E E E E E E E E E



# Test report: EN149:2001 +AI:2009

Report No: 2020 (D) - 0732 Page 8 of 10

Table -7.9.2 Penetration of filter material

Test specification: FN 140-2001 Clause 9 11

Aerosol	Condition	Sample No.	Penetration (%)	Assessment	
		11	0.379		
	As received	12	0.448		
		13 0.312			
		14	0.498		
	Sodium chloride test  Simulated wearing treatment  Mechanical strength+ Temperature conditioned  As received  Paraffin oil test  Simulated wearing treatment  Mechanical strength+ Temperature	15	0.512		
		16	0.507		
		17	0.618		
		18	0.632	]	
		19	0.668		
		20	3.71	Pass	
	As received	21	3.87		
		22	3.64		
		23	4.57		
	Simulated wearing treatment	24	4.64		
	Simulated wearing freatment	25	4.51	]	
		26	4.85	1	
	Mechanical strength+ Temperature conditioned	27	4.71	]	
		28	4.94		
Flow condition	ning: Single filter: 95.0 L/min				

# Table 7.11 Flammability Test specification: FN 149-2001 Clause 8.6

rest specifica	HOIL EIN	149-2001 Clause 8.0	
Condition	Sample No.	Result	Assessment
As received	29	Burn for 1 s	
As received	30	Burn for 1 s	Pass
Temperature conditioned	31	Burn for 1 s	rass
	32	Burn for 1 s	



## Test report: EN149:2001 +AI:2009

Report No: 2020 (D) - 0732 Page 9 of 10

Table 7.12 Carbon dioxide content of the inhalation air

Test specification: EN 149-2001 Clause 8.7

Condition	Sample No.	Resu	Assessment	
	33	0.45%		
As received	34	0.40%	Mean value 0.4%	Pass
	35	0.41%		

#### Table 7.16 Breathing resistance (mbar)

Test specification: EN 149-2001 Clause 8.9

	Elam	Flow rate		36				37				38					
As received	Flow	rate	A	В	C	D	E	A	В	C	D	Ε	A	В	C	D	E
	Inhalation	30 l/min	0.4	0.5	0.4	0.5	0.5	0.4	0.5	0.6	0.5	0.6	0.4	0.5	0.6	0.4	0.5
	innatation .	95 l/min	1.4	1.5	1.4	1.5	1.5	1.6	1.4	1.5	1.5	1.5	1.6	1.4	1.4	1.6	1.6
	Exhalation	160 l/min	1.8	1.8	1.8	1.8	1.8	1.9	1.8	1.8	1.8	1.8	1.9	1.8	1.8	1.9	1.9
	TI	200000	ĝ	27 75	39		. 3	40				41					
Simulated	Flow rate		A	В	C	D	E	A	В	C	D	Ε	A	В	C	D	Е
wearing	Inhalation	30 l/min	0.5	0.5	0.6	0.6	0.5	0.6	0.5	0.5	0.4	0.5	0.6	0.5	0.4	0.4	0.5
treatment		95 l/min	1.6	1.4	1.4	1.5	1.6	1.4	1.4	1.4	1.4	1.5	1.5	1.4	1.4	1.5	1.6
	Exhalation	160 l/min	1.9	1.8	1.8	1.9	1.9	1.7	1.7	1.7	1.7	1.8	1.9	1.8	1.7	1.9	1.9
-1	The	Flow rate		42			43				44						
т.	Flow	rate	A	В	C	D	Е	A	В	C	D	E	A	В	C	D	Е
Temperature	Tobalation	30 l/min	0.5	0.5	0.6	0.6	0.4	0.4	0.4	0.5	0.4	0.4	0.4	0.6	0.6	0.6	0.4
conditioned	Inhalation ·	95 l/min	1.5	1.5	1.6	1.4	1.6	1.4	1.6	1.5	1.5	1.5	1.6	1.4	1.5	1.6	1.6
	Exhalation	160 l/min	1.9	1.8	1.9	1.7	1.9	1.8	1.9	1.9	1.8	1.8	1.9	1.8	1.8	1.9	1.9
Assessment		20 90 80 18 18				100	Pas	s			8 7	8	3.2	800	900	8% )/s	1

A: facing directly ahead; B: facing vertically upwards; C: facing vertically downwards; D: lying on the left side; E: lying on the right side



PR 2019\_MASCHERINE\_PANTHER\_0121